

Archived Document

This archived document is no longer being reviewed through the CLSI Consensus Document Development Process. However, this document is technically valid as of September 2016. Because of its value to the laboratory community, it is being retained in CLSI's library.



CLINICAL AND
LABORATORY
STANDARDS
INSTITUTE®

October 2000

C29-A2

Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition

SAMPLE >

This document contains recommendations on the expression of the results of ion-selective electrode measurement of sodium and potassium ion activities in undiluted serum, plasma, or whole blood in clinical practice.

A standard for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeals Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeals, documented in the CLSI Standards Development Policies and Processes, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute
950 West Valley Road, Suite 2500
Wayne, PA 19087 USA
P: +1.610.688.0100
F: +1.610.688.0700
www.clsi.org
standard@clsi.org

ISBN 1-56238-410-4
ISSN 0273-3099

C29-A2
Vol. 20 No. 17
Replaces C29-A
Vol. 15 No. 1

Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition

Volume 20 Number 17

Paul D'Orazio, Ph.D.
W. Gregory Miller, Ph.D., Chairholder
Gary L. Myers, Ph.D., Vice-Chairholder
Basil T. Doumas, Ph.D.
John H. Eckfeldt, M.D., Ph.D.
Susan A. Evans, Ph.D.
Gary A. Graham, Ph.D., DABCC
Patrick J. Parsons, Ph.D.
Noel V. Stanton, M.S.

Abstract

Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition (CLSI document C29-A2) offers a protocol for standardizing instruments that contain direct ion-selective electrodes to give results in concentration terms that are verifiable to the reference method (flame photometry) for specimens with normal plasma water (i.e., lipids and proteins). The document describes the preparation of serum pools to carry out the procedure. Laboratories without the resources, equipment, or personnel to prepare the pools can purchase them from the National Institute of Standards and Technology (Gaithersburg, MD). It is recommended in C29-A2 that the accuracy of each new direct potentiometric instrument be verified with these standard pools, together with the flame photometer, if it is also used to report patient results.

Clinical and Laboratory Standards Institute (CLSI). *Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition*. CLSI document C29-A2 (ISBN 1-56238-410-4). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2000.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customerservice@clsi.org; Website: www.clsi.org.



CLINICAL AND
LABORATORY
STANDARDS
INSTITUTE®

Copyright ©2000 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition*. CLSI document C29-A2. Wayne, PA: Clinical and Laboratory Standards Institute; 2000.

Previous Editions:

December 1989, December 1992, March 1995

Archived:

September 2016

ISBN 1-56238-410-4
ISSN 0273-3099

Contents

Abstract	i
Committee Membership.....	iii
Active Membership.....	vii
Foreword	xv
1 Introduction.....	1
2 Scope.....	1
3 Definitions/Terminology.....	1
3.1 Abbreviations.....	2
4 Standard Precautions.....	2
5 Purpose.....	2
5.1 Whole Blood and Undiluted Plasma or Serum	3
5.2 "Normal" Specimens.....	3
5.3 Characteristics of a "Normal" Specimen	3
6 Background.....	3
6.1 Preanalytical Effects on Patient Sample Results: Influences of Specimen Choice, Collection, and Handling	3
6.2 Analytical Issues.....	4
6.3 Reporting Results.....	5
7 Standardization of Result Reporting.....	5
7.1 Materials	5
7.2 Sample Analysis and Data Reduction.....	6
7.3 Validation Protocol Using Patient Samples (Optional)	7
References.....	8
Appendix A. Specifications for a Human Serum Certified Reference Material (CRM) for the ISE Measurement of Sodium and Potassium in Blood	10
Appendix B. Estimated Number of Replicates for Demonstrating Equality of Flame and ISE Determination	13
Appendix C. An Example of Normalization of Linear Regression Data for Sodium to 140 mmol/L and Adjustment of the Slope to One and Intercept to Zero.....	14
Appendix D. Determination of Plasma Water Mass Concentration	16
Summary of Comments and Subcommittee Responses.....	17
Summary of Delegate Comments and Responses.....	18
Related NCCLS Publications.....	21

Foreword

Analysis of electrolytes in whole blood is becoming increasingly common, and instruments that perform these analyses (ion-selective electrodes that do not require diluted samples, i.e., direct ISE) are available from a number of manufacturers. In many clinical laboratories, instruments that use a diluted specimen [flame photometry or indirect ion-selective electrodes (ISEs)] are also still in use. Direct ISE results are not equivalent to those results obtained by the technology employing dilution for a number of reasons, foremost of which are those relating to the variation in plasma water. To avoid confusion, we recommend that results obtained by direct potentiometry be adjusted to resemble those obtained by procedures that measure concentration in plasma. Most instruments using direct ISEs have built-in conversion algorithms that can be utilized by putting the instrument in "flame" mode. However, it is clear from work performed at the National Institute of Standards and Technology (NIST) that a number of direct potentiometric systems from a variety of manufacturers do not give identical results when assaying identical specimens even in the "flame" mode.¹

The results of determinations of sodium and potassium ions in physiological fluids have been expressed in terms of (substance) concentration (mmol/L) for many years. The use of concentrations of both sodium and potassium and their reference intervals is firmly established in clinical interpretation and practice. Analytical systems that report concentration, such as flame atomic emission spectrometry (FAES) and ion-selective electrode systems using diluted samples, will continue to be used alongside direct ion-selective electrode determinations in the foreseeable future. Consequently, to introduce a new system of reporting results of sodium and potassium determinations by ion-selective electrodes in terms of ion activity carries significant risks of confusing clinical interpretation.^{2,3}

A more fundamental problem also exists. Ion-selective electrodes respond to the thermodynamic activity of ions in solution. By theory, they do not sense concentration which, in fluids such as plasma, is related in a complex way to activity.^{4, 2} Therefore, results should be expressed in terms of ion activity. Many practical difficulties exist with this approach, however. Because the activity of these ions cannot, at present, be determined with certainty, particularly in a fluid as complex as plasma, the accuracy of any determination of activity cannot be verified.

The convention recommended in this standard represents a pragmatic compromise that facilitates the introduction of ion-selective electrode determinations of sodium and potassium ion concentrations in whole blood or undiluted plasma into routine clinical practice, while minimizing the risk of clinical misinterpretation.

Key Words

Direct potentiometry, flame photometry, ion-selective electrode (ISE), potassium, sodium

Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition

1 Introduction

The objective of this standard is to describe a method to standardize direct ion-selective electrode analyzers for determination of sodium and potassium in blood plasma to units of concentration (mmol/L) in “normal” plasma, as reported by flame atomic emission spectrometry (flame photometry). This standardized method will allow clinical laboratories to use the same reference intervals, regardless of instrumentation or principle of the methodology.

2 Scope

This document addresses the determination of sodium and potassium in undiluted plasma in routine clinical practice using ion-selective electrodes.

3 Definitions/Terminology^a

Calibration material//Calibrator, n - A material or device of known, or assigned quantitative characteristics (e.g., concentration, activity, intensity, reactivity, responsiveness) used to adjust the output of a measurement procedure or to compare the response obtained with the response of a test specimen and/or sample.

Direct analysis, n - Measurement made directly on an undiluted specimen, e.g., whole blood, plasma, or sweat.

Flame mode, n - In the flame mode, a factor can be applied to results generated by direct ISE systems that makes the results comparable to those generated by indirect systems for patient specimens normal in protein and lipid content.

Heparinized, adj - Specimens anticoagulated with a heparin salt(s).

Indirect analysis, n - Systems that require dilution of the sample; **NOTE:** These include some ion-selective, electrode-based systems, as well as flame emission photometry and atomic absorption.

Plasma, n - The liquid part {of whole blood} remaining after the separation of the cellular elements ... in a receptacle containing an anticoagulant, or separated by continuous filtration or centrifugation of anticoagulated blood in an apheresis procedure.

Primary standard, n - A standard that is designated or widely acknowledged as having the highest metrological qualities and whose value is accepted without reference to other standards of the same quantity.

^a Some of these definitions are found in NCCLS document NRSCL8—Terminology and Definitions for Use in NCCLS Documents. For complete definitions and detailed source information, please refer to the most current edition of that document.

3.1 Abbreviations

		<u>Section^b</u>
ACD	Acid citrate dextrose	A2.1
CRM/SRM	Certified reference material/standard reference material	7.1/App. A
CV	Coefficient of variation	7.2.2
EP-BGpH/IFCC	Expert Panel on Blood Gas/pH of the IFCC	A3
FAAS	Flame atomic absorption spectrometry	A2.3
FAES	Flame atomic emission spectroscopy	A2.1(4)
FP	Flame photometer	6.2.2
HIV	Human immunodeficiency virus	A2.1
IFCC	International Federation of Clinical Chemistry	A3
ISE	Ion-selective electrode	3
NIST	National Institute of Standards and Technology	7.1
NRSCL	National Reference System for the Clinical Laboratory	A1.3

4 Standard Precautions

Because it is often impossible to know what might be infectious, all human blood specimens are to be treated as infectious and handled according to “standard precautions.” Standard precautions are new guidelines that combine the major features of “universal precautions and body substance isolation” practices. Standard precautions cover the transmission of any pathogen and thus are more comprehensive than universal precautions which are intended to apply only to transmission of blood-borne pathogens. Standard precaution and universal precaution guidelines are available from the U.S. Centers for Disease Control and Prevention (*Guideline for Isolation Precautions in Hospitals*. Infection Control and Hospital Epidemiology. CDC. 1996;Vol 17;1:53-80.), [MMWR 1987;36(suppl 2S):2S-18S] and (MMWR 1988;37:377-382, 387-388). For specific precautions for preventing the laboratory transmission of blood-borne infection from laboratory instruments and materials; and recommendations for the management of blood-borne exposure, refer to NCCLS document M29—*Protection of Laboratory Workers from Instrument Biohazards and Infectious Disease Transmitted by Blood, Body Fluids, and Tissue*.

5 Purpose

The purpose of this standard is to recommend a procedure for the standardization of reported, direct ion-selective electrode determination of sodium and potassium in serum, plasma, or whole blood.

^b Section where abbreviation first appears in text.

Related NCCLS Publications*

- C46-P Blood Gas and pH Analysis and Related Measurements; Proposed Guideline (2000).** This document provides clear definitions of the several quantities in current use, and provides a single source of information on appropriate specimen collection, preanalytical variables, calibration, and quality control for blood pH and gas analysis and related measurements.
- EP9-A Method Comparison and Bias Estimation Using Patient Samples; Approved Guideline (1995).** This guideline provides background information and procedures for characterizing the effects of interfering substances on test results.
- H1-A4 Evacuated Tubes and Additives for Blood Specimen Collection—Fourth Edition; Approved Standard (1996).** *American National Standard.* This standard contains requirements for blood collection tubes and additives including heparin, EDTA, and sodium citrate.
- H3-A4 Procedures for the Collection of Diagnostic Blood Specimens by Venipuncture—Third Edition; Approved Standard (1998).** This standard provides detailed descriptions and explanations of proper collection techniques, as well as hazards to patients from an inappropriate specimen collection by skin puncture procedures. (See related publications H18-A in this section, LA4-A3 in the Immunology and Ligand Assay section, and M29-A in the Microbiology section.)
- NRSCL8-A Terminology and Definitions For Use in NCCLS Documents; Approved Standard (1998).** This document provides standard definitions for use in NCCLS standards and guidelines, and for submitting candidate reference methods and materials to the National Reference System for the Clinical Laboratory (NRSCL).
- RS7-P Sodium; Proposed Summary of Methods and Materials Credentialed by the NRSCL Council (1988).**
- RS8-P Potassium; Proposed Summary of Methods and Materials Credentialed by the NRSCL Council (1988).**

* Proposed- and tentative-level documents are being advanced through the NCCLS consensus process; therefore, readers should refer to the most recent editions.

Explore the Latest Offerings From CLSI!

As we continue to set the global standard for quality in laboratory testing, we are adding products and programs to bring even more value to our members and customers.



By becoming a CLSI member, your laboratory will join 1,600+ other influential organizations all working together to further CLSI's efforts to improve health care outcomes. You can play an active role in raising global laboratory testing standards—in your laboratory, and around the world.

Find out which membership option is best for you at www.clsi.org/membership.



Find what your laboratory needs to succeed! CLSI U provides convenient, cost-effective continuing education and training resources to help you advance your professional development. We have a variety of easy-to-use, online educational resources that make eLearning stress-free and convenient for you and your staff.

See our current educational offerings at www.clsi.org/education.



When laboratory testing quality is critical, standards are needed and there is no time to waste. eCLIPSE™ Ultimate Access, our cloud-based online portal of the complete library of CLSI standards, makes it easy to quickly find the CLSI resources you need.

Learn more and purchase eCLIPSE at clsi.org/eCLIPSE.

For more information, visit www.clsi.org today.

SAMPLE



CLINICAL AND
LABORATORY
STANDARDS
INSTITUTE®

950 West Valley Road, Suite 2500, Wayne, PA 19087 USA

P: +1.610.688.0100 Toll Free (US): 877.447.1888 F: +1.610.688.0700

E: customerservice@clsi.org www.clsi.org

ISBN 1-56238-410-4