

19 February 2024

To: Recipients of CLSI MR15-Ed1

From: Jennifer K. Adams, MLS(ASCP), MSHA

Vice President, Standards and Quality

Subject: Correction

This notice is intended to inform users of a correction made to CLSI MR15, *Piperacillin-Tazobactam Breakpoints for Pseudomonas aeruginosa*, 1st ed. The correction is described below and shown as highlighted text in the table excerpt.

Table 1. Historical CLSI Piperacillin-Tazobactam Breakpoints Against *Pseudomonas aeruginosa*:

The 2012 minimal inhibitory concentration (MIC) for the resistant interpretive category is listed incorrectly as $\geq 28/4$. The interpretive category has been corrected to read $\geq 128/4$.

Table 1. Historical CLSI Piperacillin-Tazobactam Breakpoints Against *Pseudomonas* aeruginosa

	Interpretive Categories and Zone Diameter Breakpoints, nearest whole mm ^a			Interpretive Categories and MIC Breakpoints, µg/mL		
Year	S		R	S		R
1992	≥ 18	-	≤ 17	≤ 64/4	-	≥ 128/4
2012 ^b	≥ 21	15-20^	≤ 14	≤ 16/4	32/4-64/4^	≥ 128/4
2023	≥ 22	18-21^	≤ 17	≤ 16/4 ^c	32/4	≥ 64/4

Abbreviations: I, intermediate; MIC, minimal inhibitory concentration; R, resistant; S, susceptible. Symbol: ^ designates agents that have the potential to concentrate in urine.

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service (customerservice@clsi.org).

^a Disk content 100/10 μg.

^b Based on a piperacillin-tazobactam dosage regimen of ≥ 3.375 g given every 6 hours.

^c Based on a piperacillin-tazobactam dosage regimen of 4.5 g every 6 hours as a 30-minute or 3-hour infusion.