

2nd Edition

M54

Principles and Procedures for Detection and Culture of Fungi in Clinical Specimens

This guideline includes protocols, quality control parameters, and interpretive criteria for culturing fungi and for detecting and identifying fungi in direct examinations.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

M54-Ed2 January 2021 Replaces M54-A

Principles and Procedures for Detection and Culture of Fungi in Clinical Specimens

Nancy L. Wengenack, PhD, D(ABMM) Nathan P. Wiederhold, PharmD Marwan Mikheal Azar, MD Dorothy Baynham, MT(ASCP) Sanjib Bhattacharyya, MS, PhD, CQIA(ASQ) Trivikram Dasu, PhD, D(ABMLI) Philippe J. Dufresne, PhD, RMCCM William Lainhart, PhD, D(ABMM) Sixto M. Leal, Jr., MD, PhD Natasha Lindahl, BS, MT(ASCP) David H. Pincus, MS, RM/SM(NRCM), SM(ASCP) Gary W. Procop, MD, MS Sean X. Zhang, MD, PhD, D(ABMM)

Abstract

Clinical and Laboratory Standards Institute guideline M54—*Principles and Procedures for Detection and Culture of Fungi in Clinical Specimens* describes recommended processes for plating and examining fungal cultures as well as principles and procedures for the direct detection of fungi in clinical specimens, including criteria for performing and interpreting direct microscopic examinations. Safety considerations unique to mycology laboratories and a discussion of appropriate levels of laboratory service (eg, when to refer samples to more experienced laboratories) are highlighted. Specimen collection, transport, and processing recommendations including rejection criteria, are provided to guide the collection of high-quality specimens for direct examinations and fungal cultures. Fungal stains and interpretive criteria appropriate for detecting and characterizing fungal elements in direct microscopic examinations are emphasized as critical components for rapid detection of fungi. Descriptions of serological and antigen-based testing and molecular assays are also provided. Media selection, incubation conditions, and other growth requirements for fungal cultures are provided with suggested culture examination schedules, interpretations for growth on positive cultures, and reporting criteria.

Clinical and Laboratory Standards Institute (CLSI). *Principles and Procedures for Detection and Culture of Fungi in Clinical Specimens*. 2nd ed. CLSI guideline MI54 (ISBN 978-1-68440-098-0 [Print]; ISBN 978-1-68440-099-7 [Electronic]). Clinical and Laboratory Standards Institute, 2021.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace dutdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org.

If you or your organization is not a member and would like to become one, or to request a copy of the catalog, contact us at:

P: +1.610.688.0100 **F:** +1.610.688.0700 **E:** customerservice@clsi.org **W:** www.clsi.org



Copyright ©2021 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, derivative product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedures manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. Principles and Procedures for Detection and Culture of Fungi in Clinical Specimens. 2nd ed. CLSI guideline M54. Chinic and Laboratory Standards Institute; 2021.

Previous Edition: October 2012

M54-Ed2

.

ISBN 978-1-68440-098-0 (Print) ISBN 978-1-68440-099-7 (Electronic) ISSN 1558-6502 (Print) ISSN 2162-2914 (Electronic)

Volume 41, Number 1

Contents

Abstract	i
Committee Membership	iii
Foreword	vii
Chapter 1: Introduction	
1.1 Scope	
1.2 Standard Precautions	
1.3 Terminology	в
Chapter 2: Process for Directly Detecting Fungi in Specimens	
Chapter 3: Preexamination Activities	
3.1 Specimen Collection and Transport	
3.2 Specimen Rejection Criteria	
3.3 Specimen Types	
3.4 Specimen Processing	22
3.5 Specimen Retention and Storage	23
Chapter 4: Detection and Differentiation of Fungi	25
4.1 Characterizing Yeasts and Yeast-like Organisms in Direct Examinations	26
4.2 Characterizing Molds in Direct Examinations	64
4.3 Microscopy Procedures	69
4.4 Serodiagnostic Methods	75
4.5 Molecular Methods	86
Chapter 5: Fungal Specimen Culture Methods	
5.1 Specimen Plating, Media Selection, and Incubation	92
5.2 Incubation Requirements and Conditions	95
5.3 Culture Assessment	96
Chapter 6: Quality System Essentials	
6.1 Facilities and Safety Management.	
6.2 Supplier and Inventory Management: Levels of Laboratory and Referral Services	
6.3 Process Management: Quality Control of Media, Reagents, and Other Supplies	
Chapter 7: Conclusion	

. . .

Contents (Continued)

		110
Additional Pergurges		LL LL
Appendix A. Practical Applications of Flu	rescent Brighteners in Clinical Mycology	13
Appendix B. Histochemical Stains		130
Appendix C. Prevention and Control of N	te Infestations in Fungal Cultures	
Appendix D. Additional Isolation Media 1 From Patients With Cystic Fibrosis	r Culturing Sputum and Other Respirator	y Tract Samples
The Quality Management System Appr	ach	

. .

Foreword

The ability of the laboratorian to detect and characterize fungal elements directly from patient specimens can rapidly provide substantial information about possible pathogens. Therefore, characterization of yeasts and yeast-like organisms based on size and other microscopic characteristics is emphasized together with, where possible, differentiation of molds in direct examinations. Current antibody and antigen tests, as well as molecular assays for detection of fungi, are also described. Stains commonly used for direct examinations in the mycology laboratory are presented along with information on the utility of histopathology stains that may come to the attention of the mycologist.

Overview of Changes

This guideline replaces the previous edition of the approved guideline, M54-A, published in 2012. Several changes were made in this edition, including:

- Updating information, including current fungal taxonomy and nomenclature
- Expanding tables and consolidating text
- Adding new subchapters on antibody, antigen, and molecular testing
- Adding new figures, including new photographs

NOTE: The content of this guideline is supported by the CLSL consensus process and does not necessarily reflect the views of any single individual or organization.



Chapter 1 Introduction

This chapter includes:

- Guideline's scope and applicable exclusions
- Standard precautions information
- Terminology information, including:
 - Terms and definitions used in the guideline
 - Abbreviations and acronyms used in the guideline

Principles and Procedures for Detection and Culture of Fungi in Clinical Specimens

1 Introduction

1.1 Scope

This guideline provides recommendations to laboratories on procedures for collecting, processing, and handling fungal specimens and interpreting direct stain examinations and culture results. In addition, methods for direct or indirect detection from patient specimens, such as antigen, antibody, and molecular testing, are included. Because the relative importance of any fungus isolated from a patient specimen depends on the pathogenic potential of the fungus and the clinical setting in which it is isolated, these issues as well as factors to consider regarding the isolate's clinical significance are discussed.

Fungal taxonomy has been updated in this edition of M54. Direct molecular methods to detect fungi in clinical specimens are also discussed, such as the use of magnetic resonance to detect *Candida* spp. in blood and the use of PCR to detect *Pneumocystis jirovecii* in respiratory specimens, other fungal pathogens, and the emerging pathogen *Candida auris*. This guideline considers individualized quality control plan (IQCP) issues related to fungal media and includes a table listing the differential diagnosis of various yeasts and yeast-like organisms on direct examination. Antigen and antibody detection of fungi and the extent of identification needed to provide clinical and therapeutic guidance are also discussed.

The intended users of this guideline are laboratorians who process specimens for fungal culture, perform fungal direct microscopic examinations, and/or perform antibody, antigen, or molecular testing for fungi. Antifungal susceptibility testing methods (see CLSI documents M27,¹ M38,² M44,³ M51,⁴ M60,⁵ and M61⁶) are not discussed in this guideline. Although *Nocardia* spp. and other aerobic actinomycetes can be encountered growing on mycology media, methods for detecting these organisms are not discussed in this guideline. Additionally, definitive fungal identification from culture growth (eg, examination of cellulose tape preparation, matrixassisted laser desorption/ionization time of flight mass spectrometry [MALDI-TOF MS], or DNA sequencing) is outside this guideline's scope (see CLSI documents M58⁷ and MM18⁸).

1.2 Standard Precautions

Because it is often impossible to know what isolates or specimens might be infectious, all patient and laboratory specimens are treated as infectious and handled according to "standard precautions." Standard precautions are guidelines that combine the major features of "universal precautions and body substance isolation" practices. Standard precautions cover the transmission of all known infectious agents and thus are more comprehensive than universal precautions, which are intended to apply only to transmission of bloodborne pathogens. Published guidelines are available that discuss the daily operations of diagnostic medicine in humans and animals while encouraging a culture of safety in the laboratory.⁹ For specific precautions for preventing the laboratory transmission of all known infectious diseases, refer to CLSI document M29.¹⁰ For detailed information on biosafety practices specific to the mycology laboratory, see Subchapter 6.1.

3.3 Specimen Types

For detailed information on acceptable clinical specimen types for fungal culture, refer to Table 2.

Table 2. Specimen	Collection,	Transport, and	Processing Chart
-------------------	-------------	----------------	-------------------------

Specimen Site and/or Type	Collection Guidelines	Collection Container	Unacceptable Specimen Collection Method and/or Type	Specimen Processing Method	Notes
Abscesses and wounds	 The surface exudate should be removed with preservative-free, sterile saline or 70% ethanol. Aseptic aspiration from undrained abscesses should be performed using a needle and syringe. Tissue from the deepest part of the lesion plus the advancing margin should be included. 	Sterile tube	Swabs	< 2 mL direct inoculation; > 2 mL concentrate. Excessively mucoid specimens can be processed with a mucolytic agent before plating. ^a	 If grains or granules are present: The color should be documented. They should be removed and washed in sterile saline. A granule should be crushed be tween two glass slides and examined microscopically. If fungal elements are present, individual sterile, washed granules should be processed by crushing and plating directly.
Blood	 The skin should be thoroughly deansed using standard blood culture collection protocols. The maximum amount of blood recommended for the collection container being used should be collected. 	 A special mycological blood culture medium bottle or lysis centrifugation tube should be used.¹⁸⁻²⁴ Routine blood culture bottles may be appropriate for some yeasts. 	 Expired media bottles or tubes Blood cultures should not be refrigerated after specimen collection. 	The manufacturer's instructions should be followed.	Direct examination should not be performed. Refer to CLSI document M47. ²⁵

. . . .

.

• • •

4.1.1 Small-sized Yeasts (2 to 5 µm)

The yeasts and yeast-like organisms in the small-sized category that may be seen by direct examination include but are not limited to *C. auris, C. glabrata, H. capsulatum, P. jirovecii, Malassezia* spp., and *S. schenckii* species complex. The nonbudding endospores of *Coccidioides* spp., which may appear individually when the spherule has ruptured, are also in this size range. A fungus encountered primarily in Southeast Asia that also has forms in this size range is *T. marneffei*.

• *C. auris: C. auris* is similar in size to *C. glabrata* but can be morphologically distinguished from other largersized *Candida* spp. by its size and inability to form filaments *in vitro* (see Figure 2). *C. auris* very rarely forms pseudohyphae *in vivo*.¹⁴



Figure 2. Gram Stain of a C. auris Culture. Arrows point to narrow-budding yeast with no pseudohyphae.

Related CLSI Reference Materials^a

- **EP23™** Laboratory Quality Control Based on Risk Management. 1st ed., 2011. This document provides guidance based on risk management for laboratories to develop quality control plans tailored to the particular combination of measuring system, laboratory setting, and clinical application of the test.
- M22 Quality Control for Commercially Prepared Microbiological Culture Media. 3rd ed., 2004. This document contains quality assurance procedures for manufacturers and users of prepared, ready-to-use microbiological culture media.
- **M27** Reference Method for Broth Dilution Antifungal Susceptibility Testing of Yeasts. 4th ed., 2017. This standard covers antifungal agent selection and preparation, test procedure implementation and interpretation, and quality control requirements for susceptibility testing of yeasts that cause invasive fungal infections.
- **M29 Protection of Laboratory Workers From Occupationally Acquired Infections. 4th ed., 2014.** Based on US regulations, this document provides guidance on the risk of transmission of infectious agents by aerosols, droplets, blood, and body substances in a laboratory setting; specific precautions for preventing the laboratory transmission of microbial infection from laboratory instruments and materials; and recommendations for the management of exposure to infectious agents.
- M38 Reference Method for Broth Dilution Antifungal Susceptibility Testing of Filamentous Fungi. 3rd ed., 2017. This standard includes antifungal agent selection, preparation of antifungal stock solutions and dilutions for testing, test procedure implementation and interpretation, and quality control requirements for susceptibility testing of filamentous fungi (moulds) that cause invasive and cutaneous fungal infections.
- **M44** Method for Antifungal Disk Diffusion Susceptibility Testing of Yeasts. 3rd ed., 2018. This guideline provides an established methodology for disk diffusion testing of *Candida* spp., along with recommendations for results interpretation and quality control testing.
- M47 Principles and Procedures for Blood Cultures. 1st ed., 2007. This document provides recommendations for the collection, transport, and processing of blood cultures as well as guidance for the recovery of pathogens from blood specimens taken from patients who are suspected of having bacteremia or fungemia.

M51

Method for Antifungal Disk Diffusion Susceptibility Testing of Nondermatophyte Filamentous Fungi. 1st ed., 2010. This document describes the guidelines for antifungal susceptibility testing by the disk diffusion method of nondermatophyte filamentous fungi (moulds) that cause invasive disease.

^a CLSI documents are continually reviewed and revised through the CLSI consensus process; therefore, readers should refer to the most current editions.

Related CLSI Reference Materials (Continued)

M58 Methods for the Identification of Cultured Microorganisms Using Matrix-Assisted Laser Desorption/Ionization Time-of-Flight Mass Spectrometry. 1st ed., 2017. This guideline includes performance, reporting, and quality assurance recommendations for the identification of cultured microorganisms by medical laboratory professionals using matrix-assisted laser desorption/ionization time-of-flight mass spectrometry. Recommendations for end-user verification and workflow integration are also included.

- M60 Performance Standards for Antifungal Susceptibility Testing of Yeasts. 2nd ed., 2020. This document provides updated minimal inhibitory concentration, zone diameter, and quality control tables for the Clinical and Laboratory Standards Institute antifungal susceptibility testing documents M27 and M44.
- M61 Performance Standards for Antifungal Susceptibility Testing of Filamentous Fungi. 2nd ed., 2020. This document provides minimal inhibitory concentration breakpoints and quality control tables for the Clinical and Laboratory Standards Institute antifungal susceptibility testing documents M38 and M51.
- MM18 Interpretive Criteria for Identification of Bacteria and Fungi by Targeted DNA Sequencing. 2nd ed., 2018. This guideline includes information on sequencing DNA targets of cultured isolates, provides a quantitative metric for perceiving microbial diversity and can serve as the basis to identify microorganisms. By establishing interpretive criteria for microorganism identification by targeted DNA sequencing, this guideline provides structure to laboratories that identify microorganisms for medical use.



PRINT ISBN 978-1-68440-098-0 ELECTRONIC ISBN 978-1-68440-099-7 M54-Ed2